

## INTEROFFICE COMMUNICATION


# WASHINGTON STATE PATROL

TO: Dr. Barry K. Logan, State Toxicologist

FROM: Sgt. Rod Gullberg, Breath Test Section

SUBJECT: Approval of BAC Datamaster CDM Instrument

DATE: June 30, 2004



The Breath Test Section has been evaluating the BAC Datamaster CDM (National Patent Analytical Systems, Inc., Mansfield, OH) breath test instrument for the past few months. Our evaluation is complete and the summary reports have been provided for your review. The instrument has complied with our existing Quality Assurance Procedure standards and has performed with comparable accuracy and precision to the existing Datamaster instrument in both control standard and human subject testing. The Datamaster CDM is essentially the same instrument as the current Datamaster with both having the same filter frequencies, same software and most of the same hardware components. Based on this information and the results of our evaluation, we are recommending that the BAC Datamaster CDM be identified in the Washington Administrative Code as an approved evidential breath test instrument for use in Washington State.

  
RGG:rg

Cc: Lt. R. S. Reichert, Implied Consent Section

CDM Approved 6/30/2004

